

**American Society of ExtraCorporeal Technology**  
**Standards and Guidelines**  
**for Perfusion Practice**

The American Society of ExtraCorporeal Technology (AmSECT) has created the following document based on clinical evidence and currently accepted perfusion practices. Perfusionists are the only allied healthcare professionals formally trained and educated in the field of extracorporeal science and whose scope of practice expressly includes the utilization of extracorporeal devices. The document is intended to serve as a useful guide for teams developing institution-specific protocols to improve the reliability, safety and effectiveness of extracorporeal support services.

**Goal Statement**

The goal of this project was to provide Perfusionists with a framework to guide safe and effective extracorporeal support care to their patients. AmSECT recommends that clinical teams use this document as a guide for developing institution-specific protocols for patients receiving extracorporeal support.

**Approach**

In 2011, the AmSECT Board of Directors (BOD) requested the International Consortium for Evidence-Based Perfusion (ICEBP) subcommittee to review and update the Essentials and Guidelines. In 2013, the revision was completed and adopted by the membership, and a report of this work published in the Journal of Extracorporeal Technology (J Extra Corporeal Technol. 2013 Sep;45(3):156-66). In recognition of the developing role of extracorporeal support the BOD requested that the 2013 Standards and Guidelines be updated. The ICEBP undertook this review and shared the suggested revision with the BOD and the perfusion community at AmSECT's conferences in 2014 and 2015. Based on feedback from conference attendees, the ICEBP have submitted the current revised document for BOD approval to be formally presented to the membership. These Standards and Guidelines will be reviewed and updated as necessary or as deemed appropriate by AmSECT's BOD.

Three new Standards are proposed in relation to:

- 1) Support for procedures not involving cardiopulmonary bypass (e.g. off-pump or TAVR),
- 2) Staffing, and
- 3) Timing of protamine administration and cessation of cardiotomy suction.

To increase the value of this document, we have added the following new Appendices

- 1) Appendix E, which provides regulatory documentation supporting each standard and guideline,
- 2) Appendix F, which offers an example of a checklist that may be considered by perfusion teams.

Additional changes to the document include:

- 1) All additions/revisions relating to the previously approved Standards and Guidelines are noted in Red
- 2) Any deletions to standards or guidelines are noted with a ~~striketrough~~
- 3) The order of the Standard and Guidelines has been changed to reflect a contemporary lens upon the practice of extracorporeal circulatory support

To facilitate the understanding of the Standards and Guidelines, we define important terms used through the document.

**Definitions:**

**Standard:** Practices, technology and/or conduct of care that institutions shall meet in order to fulfill the minimum requirements for cardiopulmonary bypass.

**Guideline:** A recommendation that should be considered and may assist in the development and implementation of protocols.

**Protocol:** An institution-specific written document, derived from professional standards and guidelines, which contains decision and treatment algorithms.

**Word Usage:**


**Shall:** In this document, the word shall is used to indicate a mandatory requirement.

**Should:** In this document, the word should is used to indicate a recommendation.

**Surgical Care Team:** In this document, the term surgical care team is used to indicate the group surgeon, anesthesiologist, Perfusionist, nurse and technicians.

**Disclaimer:**

**AmSECT recognizes that individual medical centers may have local policies that may supersede AmSECT's Standards and Guidelines. Likewise, AmSECT recognizes that some districts or states may have laws that supersede AmSECT's Standards and Guidelines. As a result, Perfusionists practicing within those jurisdictions should comply in all respects with those policies and laws.**



**American Society of ExtraCorporeal Technology  
Standards and Guidelines  
For Perfusion Practice  
May 2017**

- Standard 1: Development of Institutionally-based Protocols**
- Standard 2: Qualification, Competency and Support Staff\***
- Standard 3: Communication**
- Standard 4: Perfusion Record**
- Standard 5: Checklist**
- Standard 6: Safety Devices**
- Standard 7: Monitoring**
- Standard 8: Anticoagulation**
- Standard 9: Gas Exchange**
- Standard 10: Blood Flow**
- Standard 11: Blood Pressure**
- Standard 12: Protamine and Cardiomy Suction**
- Standard 13: Blood Management**
- Standard 14: Level of Readiness**
- Standard 15: Staffing**
- Standard 16: Duty Hours**
- Standard 17: Quality Assurance and Improvement**
- Standard 18: Maintenance**

**Standard 1: Development of Institutionally-based Protocols**

**Standard 1.1:** As a mechanism for applying each standard to clinical practice, an institution or service provider shall develop and implement an operating procedure (protocol) for each of the standards.

**Standard 1.2:** The protocol shall be:

- Approved by the Chairman of Cardiac Surgery, or his/her designee, Director of Perfusion or equivalent, and other relevant clinical governance committees if available.
- Reviewed and revised annually or more frequently when deemed necessary.

Guideline 1.1: Deviation from protocol may be at the discretion of the Surgical Care Team and should be documented in the perfusion record.

---

## **Standard 2: Qualification, Competency and Support Staff**

**Standard 2.1:** A Perfusionist, who is Board Certified by the American Board of Cardiovascular Perfusion or who demonstrates equivalent qualifications and competency, shall conduct cardiopulmonary bypass (CPB).<sup>1</sup>

**Standard 2.2:** Perfusionist competency shall be assessed annually to evaluate compliance with departmental protocols.

**Standard 2.3:** The Perfusionist shall attend, participate, and engage in perfusion-related continuing education courses on an annual basis.<sup>2</sup>

**Standard 2.4:** Support staff shall be available on site to assist the primary Perfusionist during CPB procedures.

**Standard 2.5:** A process to educate, train, and annually evaluate perfusion staff shall be developed and followed.

Guideline 2.1: An individual graduating from an accredited perfusion education program should complete all requirements for American Board of Cardiovascular Perfusion certification within 3 years of graduation.

Guideline 2.2: A standardized process should be developed and followed to identify, orient and educate support staff to ensure they have general knowledge of the duties performed by the Perfusionist, flow of the operation and location of primary and ancillary items required during CPB. Support staff may include a Perfusionist, nursing, technical, or non- technical staff.

---

<sup>1</sup> AmSECT recognizes that individual states may license Perfusionists based on other criteria. These laws supersede this standard.

<sup>2</sup> American Board of Cardiovascular Perfusion, [www.abcp.org/](http://www.abcp.org/) (accessed November 30, 2016)

### **Standard 3: Communication**

**Standard 3.1:** A patient-specific management plan for the cardiopulmonary bypass (CPB) procedure shall be prepared and communicated to the surgical team either during the pre- operative briefing or prior to beginning the procedure.<sup>3</sup>

**Standard 3.2:** The primary Perfusionist shall use a set handoff protocol e.g. SBAR when transitioning the management of the case to a second Perfusionist.<sup>4</sup>

**Guideline 3.1:** The use of cellular telephone technology in the operating room should be guided by the principles of ST-59 Statement on use of cell phones in the operating room, written by the American College of Surgeons.<sup>5</sup>

**Guideline 3.2:** Protocol driven communication (e.g. closed-loop), should be utilized to acknowledge verbal commands, verify the content, and reduce ambiguity.<sup>6,7,8</sup>

**Guideline 3.3:** The primary Perfusionist should participate in the post-procedure debrief with the surgical team.

---

<sup>3</sup> World Health Organization surgical safety checklist and implementation manual. World Health Organization, [http://www.who.int/patientsafety/safesurgery/ss\\_checklist/en/](http://www.who.int/patientsafety/safesurgery/ss_checklist/en/) (accessed November 30, 2016)

<sup>4</sup> The Joint Commission. Hot Topics in Health Care. Transitions of Care: The need for a more effective approach to continuing patient care. [http://www.jointcommission.org/assets/1/18/hot\\_topics\\_transitions\\_of\\_care.pdf](http://www.jointcommission.org/assets/1/18/hot_topics_transitions_of_care.pdf) (accessed 14th October 2016)

<sup>5</sup> Statement on use of cell phones in the operating room, September 2008, Volume 93, Number 9. Bulletin of the American College of Surgeons, <https://www.facs.org/~media/files/publications/bulletin/2008/2008%20september%20bulletin.ashx> (accessed November 30, 2016)

<sup>6</sup> Wadhwa RK, Parker SH, Burkhart HM, Greason KL, Neal JR, Levenick KM, Wiegmann DA, Sundt TM, 3rd. Is the "sterile cockpit" concept applicable to cardiovascular surgery critical intervals or critical events? The impact of protocol-driven communication during cardiopulmonary bypass. *J Thorac Cardiovasc Surg.* 2010;139:312-319

<sup>7</sup> Whyte S, Cartmill C, Gardezi F, Reznick R, Orser BA, Doran D, Lingard L. Uptake of a team briefing in the operating theatre: A burkean dramatic analysis. *Soc Sci Med.* 2009;69:1757-1766

<sup>8</sup> de Vries EN, Prins HA, Crolla RM, den Outer AJ, van Andel G, van Helden SH, Schlack WS, van Putten MA, Gouma DJ, Dijkgraaf MG, Smorenburg SM, Boermeester MA. Effect of a comprehensive surgical safety system on patient outcomes. *N Engl J Med.* 2010;363:1928-1937

#### **Standard 4: Perfusion Record**

**Standard 4.1:** The perfusion record (written and/or electronic) for each cardiopulmonary bypass (CPB) procedure shall be included as part of the patient's permanent medical record. The perfusion records shall be maintained and stored according to institution policy for retaining patient medical records.

**Standard 4.2:** The record shall include:

- Patient information including demographics and pre-operative risk factors (Appendix A).
- Information sufficient to accurately describe the procedure, personnel, and equipment (Appendix B).
- Patient physiological parameters documented at a frequency determined by institutional protocol (Appendix C).
- Blood gas and anticoagulation monitoring results (Appendix D).
- Signature of the Perfusionist (and all relief Perfusionists) performing the procedure.

Guideline 4.1: The perfusion record should include open text (factual) commentary including supervising physician verbal orders pertinent to the CPB procedure.

Guideline 4.2: The perfusion record should include the signatures of the physician(s) providing oversight for the CPB procedure.

Guideline 4.3: Raw data (e.g. blood flow, pressure and temperature values) contained in electronic perfusion databases should be stored for a time period in accordance with your institution's policy for retaining electronic patient medical records.

**Standard 5: Checklist**

**Standard 5.1:** The Perfusionist shall use a checklist for each cardiopulmonary bypass (CPB) procedure.<sup>9</sup>

**Standard 5.2:** Checklists shall be included as part of the patient's permanent medical record.

Guideline 5.1: The Perfusionist should use checklists in a read-verify manner where critical steps that should have been performed are confirmed.<sup>10</sup> Completion of the checklist should be performed by two people, one person being the primary Perfusionist responsible for operation of the heart lung machine during the intra-operative period.

Guideline 5.2: The Perfusionist should utilize a checklist throughout the entire peri-operative period (e.g. set-up, pre-bypass, initial onset of bypass, prior to cessation of bypass, post bypass, and/or any return to bypass).

Guideline 5.3: The Perfusionist should utilize a checklist for other ancillary perfusion services (e.g. cell salvage, intra-aortic balloon pump, extracorporeal membrane oxygenation).

---

<sup>9</sup> Haynes AB, Weiser TG, Berry WR, Lipsitz SR, Breizat AH, Dellinger EP, Herbosa T, Joseph S, Kibatala PL, Lapitan MC, Merry AF, Moorthy K, Reznick RK, Taylor B, Gawande AA; Safe Surgery Saves Lives Study Group. A surgical safety checklist to reduce morbidity and mortality in a global population. N Engl J Med. 2009 29;360(5):491-9.

<sup>10</sup> Advancing Patient Safety in the U.S. Department of Veterans Affairs. Preoperative Briefing Guide for Use in the Operating Room. Commonwealth Fund Pub. 1477, Vol 9.



## **Standard 6: Safety Devices**

**Standard 6.1:** Pressure monitoring of the arterial line, cardioplegia delivery systems and venous reservoir (when augmented venous drainage is utilized), shall be employed during cardiopulmonary bypass (CPB) procedures.

- The pressure monitor shall be either servo regulated to control the arterial/cardioplegia pump or to allow interruption to the arterial/cardioplegia flow.
- The pressure monitor shall include an audible and visual alarm.

**Standard 6.2:** A bubble detector shall be employed during CPB procedures

- The gross/macro bubble detector shall be used to control the arterial pump or to allow interruption of the arterial blood flow.
- The detector system shall include an audible and visual alarm, and be positioned according to manufacturer instructions for use to enable timely identification and action.

**Standard 6.3:** A level sensor shall be employed during CPB procedures utilizing a (hard- shell) reservoir.

- The level sensor shall be either servo regulated to control the arterial pump or to allow interruption of the arterial blood flow.
- The level sensor shall include an audible and visual alarm, ~~and be~~ positioned according to manufacturer's instructions to allow an appropriate reaction time and a safe operational volume.

**Standard 6.4:** Temperature monitoring of the arterial outflow from the oxygenator shall be employed during CPB procedures.

- The temperature sensor shall include an audible and visual alarm to prevent high arterial outlet temperatures.

**Standard 6.5:** An arterial-line filter shall be employed during CPB procedures.

**Standard 6.6:** A one-way valve in the vent line shall be employed during CPB procedures.

**Standard 6.7:** A method for retrograde flow avoidance when using a centrifugal pump shall be employed during CPB procedures.

- Examples of retrograde avoidance systems may include the following:
  - One way flow valves
  - Hard stop detent controls to prevent accidental reduction in pump speed
  - Electronically activated arterial line clamps
  - Low speed visual and audible alarm.

**Standard 6.8:** An anesthetic gas scavenge line shall be employed whenever inhalation agents are introduced into the circuit during CPB procedures.

**Standard 6.9:** Hand cranks shall be readily available during CPB procedures.

**Standard 6.10:** A back-up gas supply shall be available during CPB procedures.

**Standard 6.11:** A back-up battery supply for the CPB machine shall be available during CPB procedures.

Guideline 6.1: A ventilating gas oxygen analyzer should be employed during CPB procedures.

Guideline 6.2: A level sensor should be employed during CPB procedures utilizing a soft shell reservoir.

- The level sensor should be either servo regulated to control the arterial pump or to allow interruption of the arterial blood flow.

- The level sensor should include an audible and visual alarm, and be positioned according to manufacturer's instructions to allow an appropriate reaction time and a safe operational volume.
  - The use of an air bubble detector distal to the outlet can be used utilized as a surrogate level detector.
-

**Standard 7: Monitoring**<sup>11</sup>

**Standard 7.1:** Patient arterial blood pressure shall be monitored continually during cardiopulmonary bypass (CPB).

**Standard 7.2:** Arterial line pressure shall be monitored continually during CPB.

**Standard 7.3:** Arterial blood flow shall be monitored continually during CPB.

**Standard 7.4:** Cardioplegia dose, delivery method, line pressure (antegrade), coronary sinus pressure (retrograde) and ischemic intervals shall be monitored continually during CPB.

**Standard 7.5:** Patient and device temperatures shall be monitored continually during CPB.

- Patient (e.g. nasopharyngeal, rectal, bladder, esophageal)
- Heart lung machine (arterial, venous and cardioplegia)
- Heater cooler (H<sub>2</sub>O temperature)

**Standard 7.6:** Blood gas analyses shall be monitored continually or at regular intervals during CPB (Appendix D).

**Standard 7.7:** Hematocrit (or hemoglobin) shall be monitored continually during CPB.

**Standard 7.8:** Oxygen fraction and gas flow rates shall be monitored continually during CPB (Appendix D).

**Standard 7.9:** The percentage of venous line occlusion of the venous occluder shall be monitored continually during CPB.

---

<sup>11</sup> To be performed in conjunction with Standard 3.

**Standard 7.10:** Venous oxygen saturation shall be monitored continually during CPB.

Guideline 7.1: Carbon dioxide removal should be monitored continually during CPB.

Guideline 7.2: Arterial oxygen saturation should be monitored continually during CPB.

Guideline 7.3: The following patient pressures should be monitored during CPB

- Central venous pressure and/ or
- Pulmonary artery blood pressure

Guideline 7.4: Continuous in-line blood gas monitoring should be used during CPB.

Guideline 7.5: Cerebral oximetry should be used during CPB.

Guideline 7.6: Arterial blood flow should be monitored continually at a point in the CPB circuit where it accurately reflects the flow delivered to the patient during CPB (eg distal to intra- circuit shunts).

## **Standard 8: Anticoagulation**

**Standard 8.1:** The Perfusionist, in collaboration with the physician-in-charge, shall define the intended treatment algorithm for anticoagulation management (heparin) and an alternative algorithm for when heparin is not suitable, including acceptable ranges for ACT.

**Standard 8.2:** The Perfusionist shall work closely with the surgical care team to monitor and treat the patient's anticoagulation status before, during, and after the cardiopulmonary bypass (CPB) period.

**Guideline 8.1:** The surgical care team should determine the target activated clotting time by considering relevant factors; including variability in the measurement of activated clotting time (ACT) attributed to the device's performance characteristics.

**Guideline 8.2:** Patient-specific initial heparin dose should be determined by one of the following methods:

- Weight
- Dose Response Curve (automated or manual)
- Blood Volume
- Body Surface Area

**Guideline 8.3:** Anticoagulation monitoring should include the testing of ACT. Additional monitoring tests may include:

- Heparin level measurement (e.g. heparin/protamine titration or unfractionated heparin level)
- Partial Thromboplastin Time
- Thromboelastograph

- Thrombin Time
- Anti Xa

Guideline 8.4: Additional doses of heparin during CPB should be determined by using an ACT and/or Heparin/Protamine titration<sup>12</sup>

Guideline 8.5: Heparin reversal should be confirmed by ACT and/or heparin/protamine titration.

---

<sup>12</sup> In patients requiring longer CPB times (>2 to 3 hours), maintenance of higher and/or patient- specific heparin concentrations during CPB may be considered to reduce hemostatic system activation, reduce consumption of platelets and coagulation proteins, and to reduce blood transfusion. (Class IIb, Level of evidence B). Ferraris et al 2011

## **Standard 9: Gas Exchange**

**Standard 9.1:** Gas exchange shall be maintained during cardiopulmonary bypass (CPB) according to protocol, accounting for:

- The individual patient characteristics/risk profile
- Oxygenator type, design and instructions for use
- Blood flow, temperature and metabolic demand

**Standard 9.2:** Devices used to measure gas exchange shall be calibrated according to the manufacturer's instructions for use.

**Standard 9.3:** Blood gas analysis shall be performed and recorded according to protocol.

Guideline 9.1: Point-of-Care testing should be considered to provide accurate and timely information for blood gas analysis.<sup>13</sup>

Guideline 9.2: Oxygen delivery and consumption calculations should be utilized to evaluate and optimize gas exchange.<sup>14</sup>

- Oxygen Delivery:  $DO_2 = 10 \times CI \times CaO_2$
- Oxygen Consumption:  $VO_2 = 10 \times CI \times (CaO_2 - CvO_2)$

---

<sup>13</sup> Nichols, JH. Laboratory Medicine Practice Guidelines. Evidence-based practice for point-of-care testing. American Association for Clinical Chemistry Press. 2006. <https://www.aacc.org/~media/practice-guidelines/point-of-care-testing/poct-entire-lmpg.pdf?la=en> (accessed November 30, 2016)

<sup>14</sup> De Somer F, Mulholland JW, Bryan MR, Aloisio T, Van Nooten GJ, Ranucci M. O<sub>2</sub> and CO<sub>2</sub> production during cardiopulmonary bypass as determinants of acute kidney injury: time for a goal-directed perfusion management? Crit Care. 2011 Aug 10;15(4):R192. doi: 10.1186/cc10349.



Where:

$CaO_2$  (arterial oxygen content) =  $(Hb \times 1.36 \times SaO_2) + (0.0031 \times PaO_2)$ ,

and

$CvO_2$  (mixed venous oxygen content) =  $(Hb \times 1.36 \times SvO_2) + (0.0031 \times PvO_2)$

CI = cardiac index

HB = hemoglobin

$SaO_2$  = arterial oxygen saturation

$PaO_2$  = partial pressure of oxygen in arterial blood

$SvO_2$  = venous oxygen saturation

$PvO_2$  = partial pressure of oxygen in venous blood

**Standard 10: Blood Flow**

**Standard 10.1:** Target blood flow rates shall be determined prior to cardiopulmonary bypass (CPB) according to protocol.<sup>16</sup>

**Standard 10.2:** The Perfusionist shall work closely with the surgical care team to maintain targeted blood flow rate during CPB.

Guideline 10.1: Variance from intended and targeted blood flow should be communicated to the physician-in-charge.

Guideline 10.2: Appropriate blood flow rate should be determined by evaluation of:

- Acid base balance
  1. Base Excess
- Anesthetic level
- Arterial blood pressure
- Cerebral oximetry
- Lactate burden
- Oxygen delivery and consumption (refer to guideline 10.2 for formula)
  - Venous pO<sub>2</sub>
  - Arterial pO<sub>2</sub>
  - Hemoglobin concentration
  - Arterial oxygen saturation
- Systemic vascular resistance (SVR)
- Temperature
- Venous oxygen saturation

## **Standard 11: Blood Pressure**

**Standard 11.1:** The Perfusionist, in collaboration with the physician-in-charge, shall define and communicate the intended treatment algorithm for blood pressure management prior to cardiopulmonary bypass (CPB), including acceptable ranges for blood pressure.<sup>15</sup>

**Standard 11.2:** The Perfusionist shall work closely with the surgical care team to maintain blood pressure according to protocol during CPB.

Guideline 11.1: Variance from intended and targeted blood pressure should be documented and communicated to the physician-in-charge to allow for changes in the blood pressure management plan.

---

<sup>15</sup> In many circumstances, the physician-in-charge may direct the perfusionist to modify the intended blood pressure management to address circumstances occurring during the CPB procedure.

**Standard 12. Protamine and Cardiomy Suction.**

**Standard 12.1:** Cardiomy suction shall be discontinued at the onset of protamine administration to avoid clotting within the CPB circuit.

### **Standard 13: Blood Management**

**Standard 13.1:** The Perfusionist shall participate in efforts to minimize hemodilution and avoid unnecessary blood transfusions.<sup>16</sup>

**Standard 13.2:** The Perfusionist shall minimize the cardiopulmonary bypass (CPB) circuit size to reduce prime volume.<sup>16</sup>

**Standard 13.3:** The Perfusionist shall calculate and communicate to the surgical team prior to initiating CPB, a patient's predicted post-dilutional hemoglobin or hematocrit.

Guideline 13.1: Blood management efforts should include the following<sup>16</sup>:

- Participate in pre-operative briefings (discussions) with the surgical care team (Standard 5.1) regarding transfusion strategies and target hematocrit values.
- Participation in a multidisciplinary blood management team.
- Minimize hemodilution by:
  - Matching the size of the CPB circuit to the size of the patient
  - Autologous priming of CPB circuit, including retrograde arterial and venous antegrade priming
  - Biocompatible coating on the surface of all CPB components
  - Perioperative blood cell recovery and reinfusion after being appropriately processed.
  - CPB circuit blood salvage at the end of the procedure

Guideline 13.2: Point-of-Care hemostasis monitoring should be utilized to minimize blood

---

<sup>16</sup> Ferraris VA, et al. 2011 update to the Society of Thoracic Surgeons and the Society of Cardiovascular Anesthesiologists blood conservation clinical practice guidelines. *Ann Thorac Surg* 2011 Mar;91(3):944-82.

loss. Monitoring may include:

- International normalized ratio
- Partial thromboplastin time
- Prothrombin time
- Thrombin time
- Thromboelastography/Thromboelastometry
- Platelet count
- Platelet function analysis

**Standard 14: Level of Readiness for Procedures that may require cardiopulmonary bypass**

**Standard 14.1:** Procedures identified preoperatively to be at elevated risk of requiring conversion to cardiopulmonary bypass (CPB) shall have a protocol for transition to CPB.

**Standard 14.2:** One Perfusionist shall be assigned for each such procedure.

**Standard 14.3:** A heart-lung machine consisting of a sterile extracorporeal set-up and ancillary equipment (Ref: Appendix B) shall be readily available for the procedure.

Guideline 14.1: The level of readiness for utilizing CPB during a surgical procedure should be determined through consultation with the surgical team.

Guideline 14.2: A heart-lung machine consisting of a sterile extracorporeal set-up and ancillary equipment (Ref: Appendix B) should be readily available for emergency procedures or as part of disaster planning protocols.<sup>17</sup>

---

<sup>17</sup> Preparedness for Specific Types of Emergencies. Centers for Disease Control and Prevention (<https://emergency.cdc.gov/planning/>) (accessed November 30, 2016).

**Standard 15: Staffing and On-call**

**Guideline 15.1:** The “n+1” staffing model should be utilized at all times, where “n” equals the number of operating/procedure rooms in use at any given time at a single site.<sup>18</sup>

**Guideline 15.2:** An on-call Perfusionist should be present and clinically ready for unscheduled and emergency procedures within 60 minutes of being called.\_

---

<sup>18</sup> Generally, the minimum safe number of perfusion staff: defined as  $N + 1$ , where N equals the number of operating/procedure rooms in use at any given time at a single site. (Ref: UK Code of Practice [http://www.scps.org.uk/index.php?option=com\\_content&task=view&id=34&Itemid=40](http://www.scps.org.uk/index.php?option=com_content&task=view&id=34&Itemid=40) accessed 14<sup>th</sup> October 2016). If three operating/procedure rooms are concurrently in use then the minimum safe number of clinical perfusionists available to cover this level of activity is deemed to be four. Non-qualified staff members (e.g students or staff who have not completed training adequate to meet the requirements of the activity.) must not be included in calculating the minimum safe number of staff.



**Standard 16: Duty Hours**

**Standard 16.1:** In order for the Perfusionist to ensure proper provision of care, he/she ~~must~~ shall receive an adequate rest period between scheduled work hours.<sup>19</sup>

Guideline 16.1: The Perfusionist should receive a minimum of 8 hours of rest period for every 16-hour consecutive work period.

---

<sup>19</sup> 10.0 Tiredness and European Working Time Directive (EWTD). The Society of Clinical Perfusion Scientists of Great Britain and Ireland *and* The College of Clinical Perfusion Scientists of Great Britain and Ireland Standards of Practice Document  
[http://www.scps.org.uk/index.php?option=com\\_content&task=view&id=25&Itemid=40](http://www.scps.org.uk/index.php?option=com_content&task=view&id=25&Itemid=40) (accessed November 30, 2016)

**Standard 17: Quality Assurance and Improvement**

**Standard 17.1:** The Perfusionist shall actively participate in both institutional and departmental quality assurance and improvement programs.

Guideline 17.1: The Perfusionist should collect data concerning the conduct of perfusion via a clinical registry or database.

Guideline 17.2: The Perfusionist should use such data for quality assurance, and improvement projects.<sup>20,21</sup>

---

<sup>20</sup> Warren CS, DeFoe GR, Groom RC, Pieroni JW, Groski CS, Morse CB, Connors EM, Lataille PJ, Ross CS, Likosky DS; J Extra Corpor Technol 2011: 43(2):58-63.

<sup>21</sup> Baker RA, Newland RF, Fenton C, McDonald M, Wilcox TW, Merry AF. J Extra Corpor Technol 2012: 44(1), 26-33.

## **Standard 18: Maintenance**

**Standard 18.1:** The Perfusionist shall assure that properly maintained and functioning equipment is used in the conduct of cardiopulmonary bypass (CPB), including (but not limited to):

- Heart lung machine
  - Pumps
  - Timers
  - Pressure monitors
  - Temperature monitors
  - Low Level alarm
  - Air bubble detector(s)
  - Blood flow sensors
- Heater/Cooler
- Anesthetic vaporizer
- Oxygen Blender/Flow Meter
- Oxygen analyzer
- Ancillary Equipment
  - IABP
  - VAD device
  - Cell salvage device

**Standard 18.2:** Preventive maintenance on perfusion equipment shall be performed by appropriately trained and qualified manufacturer technicians, representatives or Bio-Medical technicians. Regularly scheduled maintenance shall be documented by the perfusion department and/or Bio-Medical engineering staff. The interval of such maintenance shall be consistent with manufacturer recommendations, applicable external accrediting agency guidelines and institutional requirements.

**Standard 18.3:** The organization shall follow a protocol for perfusion equipment failures.<sup>22</sup>

---

<sup>22</sup> New CMS & Joint Commission Regulations on Medical Equipment Maintenance: Taking the Smart Approach to Compliance. ABM Healthcare Support Services. [https://www.abm.com/documents/white-papers/hss\\_new\\_cms\\_c-](https://www.abm.com/documents/white-papers/hss_new_cms_c-)

**Standard 18.4:** Appropriate backup perfusion supplies shall be readily available.

**Standard 18.5:** The organization shall follow a protocol for acknowledging and addressing perfusion equipment notices (e.g., recalls, warnings, and advisories).

### ***Relevant Publications***

---

[suite\\_whitepaper\\_081214.pdf](#) (Accessed November 30, 2016)

American Society of Extra-Corporeal Technology. Perfusion practice survey, September, 1993. *Perfusion Life* 1994; **11**: 42–45.

American Society of Extra-Corporeal Technology. Guidelines for perfusion practice. *Perfusion Life* 1995; **12**: 20–22.

American Society of Extra-Corporeal Technology. Members accept essentials; approve revised code of ethics. *Perfusion Life* 1993; **10**: 14.

Kurusz M. Standards of practice in perfusion. *Perfusion* 1994; **9**: 211–15.

Aaron G Hill, Mark Kurusz. Perfusion Standards and Practice. *Perfusion* 1997; 12:251-255.

The Society of Clinical Perfusion Scientists of Great Britain and Ireland *and* The College of Clinical Perfusion Scientists of Great Britain and Ireland

- Standards of Practice Document.  
[http://www.scps.org.uk/index.php?option=com\\_content&task=view&id=25&Itemid=40](http://www.scps.org.uk/index.php?option=com_content&task=view&id=25&Itemid=40)  
(Accessed November 30, 2016)
- Codes of Practice Document.  
[http://www.scps.org.uk/index.php?option=com\\_content&task=view&id=34&Itemid=40](http://www.scps.org.uk/index.php?option=com_content&task=view&id=34&Itemid=40)  
(Accessed November 30, 2016)

The Australian and New Zealand College of Perfusion. Regulations and Guidelines for Perfusionists.

(<http://esvc000803.wic050u.server-web.com/Documents/ANZCP%20Regulations.pdf> Accessed November 30, 2016)

## **Appendix A: Patient information**

1. Medical Record Number
2. Patient Surname, first name
3. Demographics
  - a. Age (DOB)
  - b. Gender
  - c. Height
  - d. Weight
  - e. Body surface Area (BSA)
4. Blood Type
5. Laboratory Data
  - a. Hemoglobin/Hematocrit
  - b. Predicted Hematocrit on Bypass
  - c. White Blood Cell Count
  - d. Platelet Count
  - e. aPTT
  - f. Na
  - g. K+
  - h. BUN/CR
  - i. Glucose
  - j. Other Relevant Lab values
6. Patient Allergies
7. Planned Procedure
8. Medical History/Risk Factors (recommended)
  - a. Cardiovascular
  - b. Pulmonary
  - c. Renal
  - d. Neurologic
  - e. GI/Endocrine

## Appendix B: Information sufficient to accurately describe the procedure, personnel, and equipment

1. Date of Procedure
2. Type of Procedure
3. Perfusionist(s) Name
  - a) Detail to clearly demonstrate the Perfusionist in charge of the case at all times.
4. Surgeon(s) Name
5. Anesthesiologist(s) Name
6. Nurse (s) name
7. Operating Room Number
8. Comments/Events (recommended)
9. Equipment
  - a) Heart Lung Machine
  - b) Cell Salvage (autotransfusion) Device
  - c) Heater/Cooler

Note: Items A-C ~~are often~~ must be uniquely identified (e.g. Pump 1, 2, 3 etc.) The related serial numbers for each component (e.g. roller pumps, vaporizer, blender, etc) are documented and stored locally.
10. Disposables
  - a) Oxygenator
  - b) Cardiotomy reservoir
  - c) Tubing pack/Arterial line filter
  - d) Centrifugal pump head
  - e) Cardioplegia Delivery System
  - f) Cell Salvage (autotransfusion)
  - g) Ultrafiltration Device
  - h) Arterial Cannula
  - i) Venous Cannula
  - j) Cardioplegia Cannula
  - k) Sump/vent(s)

Note: Manufacturer, model, serial and/or lot numbers should be documented with items a-k.

**Appendix C: Patient physiological and Perfusionist practice parameters documented at a frequency determined by institutional protocol.**

1. Blood Flow Rates (RPM)
2. Arterial Blood Pressure
3. Arterial Line Pressure
4. Central Venous/Pulmonary Artery Pressure
5. Vacuum Assist Venous Return (VAVR)
  - a) VAVR pressure
  - b) Venous Inlet Pressure (VIP)
6. Arterial/Venous Blood Gases
7. Venous Oxygen Saturation
8. Patient Temperatures, including:
  - a) Patient core (at least one)
    - i. Nasopharyngeal
    - ii. Bladder
    - iii. Esophageal
    - iv. Rectal
    - v. Tympanic
  - b) Optional
    - i. Myocardium
9. CPB temperatures:
  - i. Venous return blood
  - ii. Arterial blood inflow
  - a) Optional
    - i. Water bath(s)
10. Oxygenator gases including gas flow rate and concentration(s)
11. Input fluid volumes including:
  - a) Prime
  - b) Blood Products
  - c) Asanguineous Fluids
  - d) Cardioplegic Solution
  - e) Autologous Components
12. Cardioplegia
  - a) Solution (ratio)
  - b) Route
  - c) Flow
  - d) Pressure
  - e) Temperature



f) Volume

13. Output Fluid Volumes, including:

- a) Urine output
- b) Ultrafiltrate

14. Medications and/or inhalational anesthetic agents administered via extracorporeal circuit.

## **Appendix D: Blood gas, electrolyte and anticoagulation monitoring results**

1. Blood gases
  - a) pO<sub>2</sub>
  - b) pCO<sub>2</sub>
  - c) pH
  - d) Base excess
  - e) Bicarbonate concentration
  - f) Saturation
  - g) Potassium concentration
  - h) Ionized calcium concentration
  - i) Sodium concentration
  - j) Lactate
  - k) Glucose
  - l) Hemoglobin/hematocrit
2. Activated Clotting Times (ACT) and/or Heparin/Protamine Assay Results and/or Thromboelastography Results

## Appendix E: Regulatory documents, Revision 2016

### REGULATORY CITATION LEGEND

Regulations, Standards and Guidelines Resources	Citation Prefix
AABB Standards for Perioperative Autologous Blood Collection and Administration (6 <sup>th</sup> Edition 2014)	AABB
College of American Pathologists (7/28/2015 Checklists)	CAP
Center for Improvement in Healthcare Quality (April 2016)	CIHQ
Centers for Medicare & Medicaid Conditions of Participation (CoP) – Hospitals (Title 42 Part 482)	CMS-H
CLIA Laboratory Regulations	CMS-L
Commission on Office Laboratory Accreditation (January 2016)	COLA
Healthcare Facility Accreditation Program (2015 v2)	HFAP
National Integrated Accreditation for Healthcare Organizations (Rev 11 6-17-2014)	NIAHO
International Organization for Standardization (Standard 9001:2008)	ISO 9001
Joint Commission Hospital Accreditation Standards 2016	TJC-H
Joint Commission Laboratory Accreditation Standards 2016	TJC-L

**Please note, the ISO 9001 standards are included due to the link between NIAHO Accreditation and the requirement for the hospital to become either ISO Compliant or Certified.**

<b><u>Standard/Guideline</u></b>	<i>Regulations, Standards and Guidelines Resources</i>	<b><u>Section</u></b>
<b><u>Standard 1.1</u></b>	AABB	1.3, 6.0, 6.1.1
	CAP-C	COM.10000
	CAP-G	GEN.20374, GEN.20375
	CMS-H	§482.11
	HFAP	30.00.09
	NIAHO	QM.1_SR.1a(2); QM.3; GB.1_SR.1a; SS.1
	ISO 9001	4.1; 4.2; 4.2.1; 4.2.2; 5.1
	TJC-HAP	LD.04.01.07; LD.04.01.07_EP2; LD.04.04.07_EP1-EP3; NS.02.02.01_EP3; NS.02.03.01
	TJC-L	DC.01.01.01_EP1-EP3; DC.02.02.01_EP1-EP4
<b>Standard 1.2</b> • <u>Dot point 1</u>	AABB	1.1.1; 1.3; 1.4; 6.0; 6.1 (6.1.1, 6.1.3)
	CAP-C	COM.10000; COM.10200
	CAP-G	GEN.20375
	CIHQ	GL-4
	CMS – L	§493.1200 (a-c)
	COLA	ORG 11 E; ORG 12 R; LDR 3 E; LDR 5 E
	HFAP	30.00.09
	NIAHO	NS.2_SR.3
	ISO 9001	4.2.3
	TJC-HAP	LD.04.01.07_EP1; LD.04.04.07_EP4; NR.02.03.01_EP1-EP2;

	<i>TJC-L</i>	<i>DC.02.01.01</i>
• <u>Dot point 2</u>	<i>AABB</i>	<i>6.1.4 (biennial)</i>
	<i>CAP-C</i>	<i>COM.10100 (biennial)</i>
	<i>CIHQ</i>	<i>GL-4 (triennial)</i>
	<i>COLA</i>	<i>ORG 15 R (annual)</i>
	<i>NIAHO</i>	<i>QM.5 (annual), SM.3_SR.6</i>
	<i>ISO 9001</i>	<i>4.2.3, 5.6.1</i>
<b><u>Guideline 1.1</u></b>		
	<i>AABB</i>	<i>1.3.1, 5.4.2.2.1</i>
	<i>CAP-C</i>	<i>COM.10000</i>
	<i>NIAHO</i>	<i>QM.5</i>
	<i>ISO-9001</i>	<i>1.2</i>
<b><u>Standard 2.1</u></b>		
	<i>AABB</i>	<i>2.1; 2.1.1; 2.1.3</i>
	<i>CAP-G</i>	<i>GEN.54400, GEN.54750, GEN.55500</i>
	<i>CAP-P</i>	<i>POC.06800</i>
	<i>CIHQ</i>	<i>GL-3(G), HR-3(C), HR-4(E), MS-3(E), MS-5(B)</i>
	<i>CMS-H</i>	<i>§482.11(c), §482.23(3), §482.23(5), §482.51(4)</i>
	<i>CMS-L</i>	<i>§493.1423(e), §493.1423</i>
	<i>COLA</i>	<i>PER 2 E, PER 3 R, QC 31</i>
	<i>HFAP</i>	<i>01.00.04, 03.00.02, 03.01.06, 15.02.39, 16.00.04, 16.00.11, 18.00.06, 30.00.05,</i>
	<i>NIAHO</i>	<i>GB.1_SR.1c, NS.1, SM.1, SM.2, SS.3_SR.1</i>
	<i>ISO 9001</i>	<i>6.2.1, 6.2.2</i>
	<i>TJC-HAP</i>	<i>HR.01.02.01, HR.01.02.05, HR.01.06.01</i>
	<i>TJC-L</i>	<i>DC.02.02.01_EP1, HR.01.02.05_EP1-EP3, EP6, HR.01.02.07_EP1-EP2</i>

<b><u>Standard 2.2</u></b>		
	AABB	2.1.3, 2.1.3.1
	CAP-G	GEN.55500, GEN.57000
	CAP-P	POC.06910
	CIHQ	HR-3(C)
	CMS-H	§482.23(3)
	CMS-L	§493.1235, §493.1423
	COLA	PER 5 R, QC 31
	NIAHO	SM.7_SR.1, SM.7_SR.2, SS.3_SR.1
	TJC-HAP	HR.01.06.01, HR.01.07.01 (EP1, EP2, EP5)
	TJC-L	HR.01.07.01_EP1-EP2
<b><u>Standard 2.3</u></b>		
	AABB	2.1.4
	CAP-G	GEN.54200
	CIHQ	MS-3(E)
	CMS-L	§493.557(a)(3)(iii)
	COLA	PER 6 R
	HFAP	01.00.04, 03.00.02, 16.00.06
	NIAHO	MS.10, SM.7_SR.6
	ISO 9100	6.2.2(e)
	TJC-HAP	HR.01.05.03
	TJC-L	HR.01.05.03_EP1, EP4-EP7
<b><u>Standard 2.5</u></b>		
	AABB	2.1.1, 2.1.2, 2.1.3, 2.1.4
	CAP-G	GEN.54200, GEN.54400, GEN.54750, GEN.55500, GEN.57000
	CIHQ	GL-3(G), HR-3(C), HR-4(E), MS-3(E), MS-5(B)

	<i>CMS-L</i>	<i>§493.1423(e), §493.1423, §493.1235, §493.1423, §493.557(a)(3)(iii)</i>
	<i>COLA</i>	<i>PER 2 E, PER 3 R, PER 5 R, QC 31</i>
	<i>HFAP</i>	<i>01.00.04, 03.00.02, 03.01.06, 15.02.39, 16.00.04, 16.00.11, 18.00.06</i>
	<i>NIAHO</i>	<i>GB.1_SR.1c</i>
	<i>ISO 9001</i>	<i>6.2.1 (Note), 6.2.2</i>
	<i>TJC-HAP</i>	<i>HR.01.05.03_EP1, EP4</i>
	<i>TJC-L</i>	<i>HR.01.05.03_EP1, EP4-EP7</i>

<b><u>Guideline 2.2</u></b>		
	<i>CMS-H</i>	§482.51(3)
	<i>HFAP</i>	18.00.07, 30.00.04
	<i>NIAHO</i>	SS.2_SR.3
	<i>ISO 9001</i>	6.2.1 (Note), 6.2.2(d)
<b><u>Standard 3.1</u></b>		
	<i>CIHQ</i>	NS-3
	<i>CMS-H</i>	§482.23(b)(4)
	<i>HFAP</i>	10.00.03; 10.01.26; 10.01.28; 16.00.10; 26.00.08; 26.0.11; 27.01.18
	<i>NIAHO</i>	NS.3_SR.1
	<i>TJC- HAP</i>	PC.01.03.01_EP1, EP3; PC.02.02.01_EP1; PC.02.02.01_EP1-EP2; UP.01.03.01_EP1-EP5
<b><u>Standard 3.2</u></b>		
	<i>TJC- HAP</i>	PC.02.02.01_EP1-EP2
	<i>TJC-L</i>	DC.03.03.01_EP1
<b><u>Guideline 3.2</u></b>		
	<i>AABB</i>	5.2.3
	<i>HFAP</i>	16.01.03, 16.01.04, 16.01.05
	<i>NIAHO</i>	MM.4_SR.2-SR.4
	<i>TJC- HAP</i>	LD.03.04.01_EP1; LD.03.04.02_EP3; LD.03.04.01_EP5
<b><u>Standard 4.1</u></b>	<i>AABB</i>	5.1.6.1; 6.2; 6.2.1
	<i>CAP-G</i>	GEN.20377
	<i>CAP-P</i>	POC.04400
	<i>CIHQ</i>	MR-4; OI-8; AN-2
	<i>CMS-H</i>	§482.24
	<i>HFAP</i>	10.00.03; 10.01.01; 10.01.02;



	<i>NIAHO</i>	<i>SS.6; AN.3; MR.2; MR.3_SR.1; MR.5; MR.7</i>
	<i>ISO 9001</i>	<i>4.2.1(c), 4.2.1(d)</i>
	<i>TJC-H</i>	<i>RC.01.01.01_EP1; RC.01.05.01</i>
<b><u>Standard 4.2</u></b>		
<b><u>Dot point 1, Appendix A</u></b>	<i>AABB</i>	<i>6.2; 6.2.1</i>
	<i>CAP-P</i>	<i>POC.04400</i>
	<i>CIHQ</i>	<i>OI-7; OI-8; AN-2</i>
	<i>CMS-H</i>	<i>§482.24</i>
	<i>HFAP</i>	<i>30.00.18</i>
	<i>NIAHO</i>	<i>SS.6; MR.5</i>
	<i>TJC-H</i>	<i>RC.01.01.01_EP5</i>
<b><u>Dot point 2, Appendix B</u></b>	<i>AABB</i>	<i>6.2.4</i>
	<i>CIHQ</i>	<i>OI-7</i>
	<i>CMS-H</i>	<i>§482.51</i>
	<i>HFAP</i>	<i>10.01.03; 30.00.18</i>
	<i>NIAHO</i>	<i>SS.6; SS.8 (SR.1 - SR.3); AN.3 (SR.2c, SR.2d1); MR.5; MR.7</i>
	<i>TJC-H</i>	<i>RC.01.01.01; RC.02.01.01</i>
<b><u>Dot point 3, Appendix C</u></b>	<i>AABB</i>	<i>6.2.4</i>
	<i>CIHQ</i>	<i>AN-2</i>
	<i>CMS-H</i>	<i>§482.24; §482.52</i>
	<i>HFAP</i>	<i>0.01.03; 30.00.19</i>
	<i>NIAHO</i>	<i>SS.6; SS.8 (SR.1 – SR.3); AN.3 (SR.2c, SR.2d1); MR.5_SR.1c; MR.7</i>
	<i>TJC-H</i>	<i>RC.01.01.01_EP7</i>

<b><u>Dot point 4, Appendix D</u></b>	CAP-C	COM.29950
	CIHQ	AN-2
	CMS-H	§482.24
	HFAP	10.01.03; 30.00.19
	NIAHO	SS.6; SS.8 (SR.1 - SR.3); AN.3 (SR.2c, SR.2d1); MR.5_SR.1c; MR.7
	TJC-H	RC.01.01.01_EP7
<b><u>Dot point 5</u></b>		
	AABB	6.2.4
	CAP-P	POC.04700
	CIHQ	MR-4
	CMS-H	§482.23; §482.24; §482.51
	HFAP	10.01.03; 10.01.04; 30.00.19
	NIAHO	SS.8_SR.2; MR.5 (SR.2b, SR.4, SR.4a); MR.6
	TJC-H	RC.01.02.01; RC.02.03.07_EP1
<b><u>Guideline 4.1</u></b>	NIAHO	MR.5 (SR.2 – SR.5)
<b><u>Guideline 4.2</u></b>	AABB	5.2.3
	CIQH	MR-4
	CMS-H	§482.23; §482.24; §482.51
	COLA	WAV 9 R
	HFAP	10.01.03; 10.01.04; 30.00.19
	NIAHO	MR.5 (SR.2b, SR.3, SR.4, SR.5)
	TJC-H	RC.01.02.01; RC.02.03.07
<b><u>Guideline 4.3</u></b>	AABB	6.2.8; 6.2.9
	CAP-G	GEN.20377; 20425
	CIHQ	MR-3
	CMS-H	§482.23; §482.24

	<i>CMS-L</i>	§493.1101; §493.1105
	<i>COLA</i>	<i>WAV 9 R</i>
	<i>HFAP</i>	<i>10.00.03</i>
	<i>NIAHO</i>	<i>MR.3 (SR.1 – SR.2)</i>
	<i>TJC-H</i>	<i>RC.01.05.01</i>
	<i>TJC-L</i>	<i>DC.02.04.01</i>
<b><u>Standard 5.1</u></b>	<i>TJC-H</i>	<i>UP.01.01.01</i>
<b><u>Standard 6</u></b>	<i>NIAHO</i>	<i>SS.1; AS.1</i>
	<i>TJC-H</i>	<i>NPSG.06.01.01; LD.04.04.05</i>
<b><u>Standard 6.1</u></b>	<i>CIQH</i>	<i>QS-9</i>
	<i>TJC-H</i>	<i>NPSG.06.01.01</i>
<b><u>Standard 6.2</u></b>	<i>CIQH</i>	<i>QS-9</i>
	<i>TJC-H</i>	<i>NPSG.06.01.01</i>
<b><u>Standard 6.3</u></b>	<i>CIQH</i>	<i>QS-9</i>
	<i>TJC-H</i>	<i>NPSG.06.01.01</i>
<b><u>Standard 6.4</u></b>	<i>CIQH</i>	<i>QS-9</i>
	<i>TJC-H</i>	<i>NPSG.06.01.01</i>
<b><u>Standard 6.7</u></b>	<i>CIQH</i>	<i>QS-9</i>
	<i>TJC-H</i>	<i>NPSG.06.01.01</i>
<b><u>Guideline 6.2</u></b>	<i>CIQH</i>	<i>QS-9</i>
	<i>TJC-H</i>	<i>NPSG.06.01.01</i>
<b><u>Standard 7</u></b>	<i>CIHQ</i>	<i>AN-2 E</i>
	<i>HFAP</i>	<i>15.02.17</i>
	<i>NIAHO</i>	<i>AS.3_SR.2d(1)</i>
	<i>TJC-H</i>	<i>PC.01.02.01</i>
<b><u>Standard 8</u></b>	<i>TJC</i>	<i>NPSG.03.05.01</i>

<b><u>Standard 8.1</u></b>	CIHQ	NS-3
	CMS-H	§482.23(b)(4)
	HFAP	10.00.03; 10.01.26; 10.01.28; 16.00.10; 26.00.08; 26.0.11; 27.01.18
	NIAHO	NS.3_SR.1
	TJC- HAP	PC.01.03.01_EP1, EP3; PC.02.01.01_EP1; PC.02.02.01_EP1-EP2; UP01.03.01_EP1-EP5
<b><u>Standard 9.2</u></b>	CAP-C	COM.40610;
	CAP-P	POC.07300; POC.07512; POC.07540; POC08980; POC.09035; POC.09090; POC09145
	COLA	LDR 2 E; QC 1 E; CA 1 R
	TJC-L	EC.02.04.03; QSA.02.02.01; QSA.02.03.01
<b><u>Standard 9.3</u></b>	CAP-G	GEN.41304;
	CAP-P	POC.04400; POC.04700
	COLA	LIS 2.7; APM 18 (PST) R
	TJC-L	DC.02.03.01
<b><u>Guideline 9.1</u></b>	CAP-G	GEN.41304; GEN.41345
	TJC-L	QSA.02.10.01; QSA.06.01.01; DC.02.03.01
<b><u>Standard 10.1</u></b>	CIHQ	NS-3
	CMS-H	§482.23(b)(4)
	HFAP	10.00.03; 10.01.26; 10.01.28; 16.00.10; 26.00.08; 26.0.11; 27.01.18
	NIAHO	NS.3_SR.1
	TJC- HAP	PC.01.03.01_EP1, EP3; PC.02.01.01_EP1; PC.02.02.01_EP1-EP2; UP.01.03.01_EP1-EP5
<b><u>Standard 11.1</u></b>	CIHQ	NS-3
	CMS-H	§482.23(b)(4)

	<i>HFAP</i>	10.00.03; 10.01.26; 10.01.28; 16.00.10; 26.00.08; 26.0.11; 27.01.18
	<i>NIAHO</i>	NS.3_SR.1
	<i>TJC- HAP</i>	PC.01.03.01_EP1, EP3; PC.02.01.01_EP1; PC.02.02.01_EP1-EP2; UP.01.03.01_EP1-EP5
<b><u>Standard 11.2</u></b>	<i>CIHQ</i>	NS-3
	<i>CMS-H</i>	§482.23(b)(4)
	<i>HFAP</i>	10.00.03; 10.01.26; 10.01.28; 16.00.10; 26.00.08; 26.0.11; 27.01.18
	<i>NIAHO</i>	NS.3_SR.1
	<i>TJC- HAP</i>	PC.01.03.01_EP1, EP3; PC.02.01.01_EP1; PC.02.02.01_EP1-EP2; UP.01.03.01_EP1-EP5
<b><u>Guideline 11.1</u></b>	<i>AABB</i>	1.3.1; 5.4.2.2.1
	<i>CAP-C</i>	COM.10000
	<i>NIAHO</i>	QM.5
	<i>ISO-9001</i>	1.2
<b><u>Standard 12.1</u></b>	<i>AABB</i>	5.2.3
	<i>HFAP</i>	16.01.03; 16.01.04; 16.01.05
	<i>NIAHO</i>	MM.4_SR.2-SR.4;
	<i>TJC- HAP</i>	LD.03.04.01_EP1; LD.03.04.02_EP3; LD.03.04.01_EP5
<b><u>Standard 14.1</u></b>	<i>CIHQ</i>	NS-3
	<i>CMS-H</i>	§482.23(b)(4)
	<i>HFAP</i>	10.00.03; 10.01.26; 10.01.28; 16.00.10; 26.00.08; 26.0.11; 27.01.18
	<i>NIAHO</i>	NS.3_SR.1
	<i>TJC- HAP</i>	PC.01.03.01_EP1,EP3; PC.02.01.01_EP1; PC.02.02.01_EP1-EP2; UP.01.03.01_EP1-EP5

<b><u>Guideline 14.1</u></b>	CIHQ	NS-3
	CMS-H	§482.23(b)(4)
	HFAP	10.00.03; 10.01.26; 10.01.28; 16.00.10; 26.00.08; 26.0.11; 27.01.18
	NIAHO	NS.3_SR.1
	TJC- HAP	PC.01.03.01_EP1, EP3; PC.02.01.01_EP1; PC.02.02.01_EP1-EP2; UP.01.03.01_EP1-EP5
<b><u>Standard 17.1</u></b>	AABB	5.1.2; 8.2; 9.0
	CAP-C	COM.04000; COM.04200
	CAP-G	GEN.13806
	CIHQ	QA-1
	CMS-H	§482.21
	CMS-L	§493.1200; §493.1230; §493.1239
	COLA	QA 1 E
	HFAP	12.00.00; 12.00.04
	NIAHO	QM.1 (SR.1-SR.2); QM.2; QM.3; QM.6
	ISO 9001	8.1; 8.2.1; 8.5.1; 8.5.3
	TJC-H	LD.04.04.01 (EP1-EP4); PI.01.01.01 (EP1-EP3)
	TJC-L	PI.01.01.01
<b><u>Guideline 17.1</u></b>	AABB	5.1.2.1; 5.1.2.2; 8.3; 9.0; 9.1
	CAP-C	COM.04200
	CAP-G	GEN.20316
	CIHQ	QA-2 (A-C)
	CMS-H	§482.21
	COLA	QA 2 E
	HFAP	12.00.01; 12.00.04
	NIAHO	QM.5; QM.7
	ISO 9001	8.2.3

	TJC-H	PI.01.01.01 (EP
	TJC-L	PI.02.01.01
<b><u>Guideline 17.2</u></b>	AABB	5.1.2.1; 5.1.2.2; 8.3; 9.0; 9.1; 9.2
	CAP-G	GEN.16902; GEN.20316
	CIQH	QA-2 (D-E); QA-4; QA-5
	CMS-H	§482.21
	CMS-L	§493.1200; §493.1230; §493.1239
	COLA	QA 3 R; QA 4 R; QA 5 R
	HFAP	12.00.02; 12.00.04; 12.01.02
	NIAHO	QM.7; QM.8
	ISO 9001	8.2.2; 8.3; 8.4; 8.5.1; 8.5.2; 8.5.3
	TJC-H	PI.02.01.01; PI.03.01.01
	TJC-L	PI.03.01.01
<b><u>Standard 18.1</u></b>	AABB	3.5; 3.5.1; 3.5.1.1
	CIHQ	CE-8_A
	CMS-H	§482.26; §482.41; §482.53
	CMS-L	§493.1101; §493.1254
	HFAP	11.06.09; 11.06.10
	NIAHO	PE.1; PE.7
	TJC-H	EC.02.04.01; EC.02.04.03
	TJC-L	EC.02.04.01; EC.02.04.03
<b><u>Standard 18.2</u></b>	AABB	3.5; 3.5.1; 3.5.1.1
	CIHQ	CE-8 (B, D)
	CMS-H	§482.26; §482.41; §482.53
	CMS-L	§493.1101; §493.1254
	HFAP	11.06.09
	NIAHO	PE.1; PE.7_SR.6
	TJC-H	EC.02.04.01; EC.02.04.03

	<i>TJC-L</i>	<i>EC.02.04.01; EC.02.04.03</i>
<b><u>Standard 18.3</u></b>	<i>CIQH</i>	<i>CE-8 (M, N)</i>
	<i>NIAHO</i>	<i>PE.7 (SR.4-SR.5)</i>
	<i>TJC-H</i>	<i>EC.02.04.01_EP9</i>
	<i>TJC-L</i>	<i>EC.02.04.01; EC.02.04.03</i>
<b><u>Standard 18.4</u></b>	<i>NIAHO</i>	<i>PE.7</i>
<b><u>Standard 18.5</u></b>	<i>HFAP</i>	<i>08.00.06; 25.00.00</i>
	<i>CMS-H</i>	<i>§482.25</i>
	<i>NIAHO</i>	<i>PE.1; PE.3; PE.7</i>
	<i>TJC-H</i>	<i>EC.02.02.01_EP11; MM.05.01.017</i>
	<i>TJC-L</i>	<i>EC.02.02.01_EP11</i>



## Appendix F: Perfusion Checklist

### Perfusion Checklist

Patient ID \_\_\_\_\_

Check each item when completed, sign and date. If not applicable, ~~draw line through~~. ***Bold italicized items for expedited set-up.***

- **PATIENT**
  - Patient identity confirmed***
  - Procedure confirmed***
  - Blood type, antibodies confirmed***
  - Allergies checked***
  - Blood bank number confirmed
  - Medical record number confirmed
  - Chart reviewed
- **STERILITY/CLEANLINESS**
  - Components checked for package integrity/expiration***
  - Equipment clean
  - Heat exchanger(s) leak-tested
- **PUMP**
  - Occlusion(s) set***
  - Speed controls operational***
  - Flow meter in correct direction and calibration***
  - Flow rate indicator correct for patient and/or tubing size***
  - Rollers rotate freely***
  - Pump head rotation smooth and quiet
  - Holders secure
  - Servoregulated connections tested
- **ELECTRICAL**
  - Power cord(s) connection(s) secure***
  - Servoregulation connections secure***
  - Batteries charged and operational
- **CARDIOPLEGIA**
  - System debubbled and operational***

- System leak-free after pressurization
- Solution(s) checked
- **GAS SUPPLY**
  - Gas line(s) and filler connections secure*
  - Gas exhaust unobstructed*
  - Source and appropriate connections of gas(es) confirmed*
  - Flow meter / gas blender operational*
  - Hoses leak-free
  - Anesthetic gas scavenge line operational
- **COMPONENTS**
  - System debubbled and operational*
  - Connections / stopcocks / caps secure*
  - Appropriate lines claimed / shunts closed*
  - Tubing direction traced and correct*
  - Patency of arterial line / cannula confirmed*
  - No tubing kinks noted
  - One-way valve(s) in correct direction
  - Leak-free after pressurization
- **SAFETY MECHANISMS**
  - Alarms operational, audible and engaged*
  - Arterial filter / bubble trap debubbled*
  - Cardiotomy / hardshell venous reservoir(s) vented*
  - Vent(s) tested*
  - Venous line occluder(s) calibrated and tested
  - Devices securely attached to console
- **ASSISTED VENOUS RETURN**
  - Cardiotomy positive-pressure relief valve present*
  - Negative- pressure relief valve unobstructed*
  - Vacuum regulator operational
- **MONITORING**
  - Circuit / patient temperature probes placed*
  - Pressure transducers / monitors calibrated and on proper scales
  - Inline sensors calibrated
  - Oxygen analyzer calibrated

- **ANTICOAGULATION**
  - Heparin time and dose confirmed*
  - Anticoagulation tested and reported
- **TEMPERATURE CONTROL**
  - Water source(s) connected and operational*
  - Temperature range(s) tested and operational
  - Water lines unobstructed
- **SUPPLIES**
  - Tubing clamps available*
  - Drugs available and properly labeled
  - Solutions available
  - Blood products available
  - Sampling syringes / laboratory tubes available
  - Anesthetic vaporizer correct
  - Vaporizer operational and filled
- **BACKUP**
  - Hand cranks available*
  - Duplicate circuit components / hardware available*
  - Emergency lighting / flashlight available
  - Backup full oxygen tank with flow meter available
  - Ice available
- **EMERGENT RESTART OF BYPASS**
  - Heparin time and dose confirmed*
  - Components debubbled*
  - Gas flow confirmed*
  - Alarms reengaged*
  - Water source(s) connected
- **TERMINATION CHECKLIST**
  - Venous assist off / cardiotomy / venous reservoirs vented*
  - Shunt(s) closed*
  - Vent(s) clamed / removed*
- **POSTBYPASS CHECKLIST**
  - Announce bypass terminated*
  - Arterial and venous lines clamped*

- Arterial circuit bubble-free before transfusing perfusate**
- Pump suction(s) off

Comments:

Signature: \_\_\_\_\_  
Date: \_\_\_\_\_ Time: \_\_\_\_\_

---

**These perfusion checklists, or a reasonable equivalent, should be used in perfusion practice. This is a guideline, which tPerfusionists are encouraged to modify to accommodate difference in circuit design and variations in institutional clinical practice. Users should refer to manufacturers' information, including Instructions for Use, for specific procedures and/or precautions. AmSECT disclaims any and all liability and responsibility for injury and damages resulting from following this suggested checklist. Origination 1990; revision 2004 by AmSECT Quality Committee.**